

Laboratory Accreditation Programmes

Schedule to CERTIFICATE OF ACCREDITATION	
Laboratory	Callaghan Innovation Measurement Standards Laboratory of New Zealand
Address	PO Box 31310, Lower Hutt, 5040 69 Gracefield Road, Gracefield, Lower Hutt, 5010
Telephone	04 931-3000
Fax	04 931-3117
URL	http://www.measurement.govt.nz/
Authorised Representative	Dr Blair Hall Quality Manager
Client No.	8
Programme	Metrology & Calibration Laboratory
Accreditation Number	1
Initial Accreditation Date	30 July 2004
Conformance Standard	NZS ISO/IEC 17025:2005 General requirements for the competence of testing and calibration laboratories
Testing Services Summary	5.01 Engineers' Limit Gauges 5.02 Jigs, Fixtures, Cutting Tools and Components 5.04 Machine Tools 5.05 Geometric Form 5.11 Working Standards of Length and Angle 5.12 Precision Measuring Instruments 5.14 Laser Frequency 5.21 Masses 5.31 Volumetric Equipment 5.32 Density 5.33 Hydrometers 5.35 Hygrometry 5.41 Barometers 5.42 Differential Pressure Measuring Devices (including Manometers) 5.43 Pressure Gauge Testers and Pressure Balances 5.44 Pressure and Vacuum Measurement 5.61 Temperature Measuring Equipment 5.65 Photometers and Radiometers 5.66 Lamps, LEDs, Lasers and Other Light Sources 5.67 Colour of Light Sources and Colorimeters 5.68 Optical Properties of Materials: Spectral 5.69 Optical Properties of Materials: Spectrally integrated 5.70 Optical Instruments

Authorised: General Manager		Issue 46	Date: 28/11/17	Page 1 of 26
--------------------------------	---	----------	----------------	--------------

Laboratory Accreditation Programmes

Schedule to CERTIFICATE OF ACCREDITATION																																																			
	5.82 Resistors, Resistance Boxes and Potential Dividers 5.84 Capacitors 5.85 Inductors and Transformers 5.86 Voltage Standards and Current Standards 5.87 Transfer Instruments (AC/DC) 5.88 Calibrators for Instrumentation 5.89 Indicating Instruments and Recording Instruments 5.90 Bridges, Potentiometers and Test Sets 5.91 Frequency Measurement and Time Measurement 5.92 Waveform Measurement 5.93 Signal Sources 5.95 Communications Equipment 5.97 High Voltage Testing																																																		
Signatories	<table border="0"> <tr> <td>Dr Tim Armstrong</td> <td>5.91, 5.92(a)(c), 5.93(a)</td> </tr> <tr> <td>Dr Laurie Christian</td> <td>5.82, 5.85(d), 5.86, 5.87, 5.88, 5.89, 5.90, 5.92(b), 5.93(b), 5.97</td> </tr> <tr> <td>Dr Mark Clarkson</td> <td>5.41, 5.42, 5.43, 5.44</td> </tr> <tr> <td>Mr David Cochrane</td> <td>5.05(d)(ii), 5.70</td> </tr> <tr> <td>Dr Adam Dunford</td> <td>5.91(a)(c)(d)(g), 5.92(a)(c), 5.93(a)</td> </tr> <tr> <td>Dr Murray Early</td> <td>5.82, 5.86, 5.87, 5.88, 5.89(a)(b)(c)(d)(i), 5.90(a)(f)(g), 5.92(b), 5.93(b), 5.97</td> </tr> <tr> <td>Mr Hamish Edgar</td> <td>5.61</td> </tr> <tr> <td>Dr Lucy Forde</td> <td>5.02, 5.04, 5.11(f)(i)(n), 5.12, 5.14</td> </tr> <tr> <td>Dr Blair Hall</td> <td>5.93(b), 5.95</td> </tr> <tr> <td>Mr John Hamlin</td> <td>5.65, 5.66, 5.67, 5.68, 5.69</td> </tr> <tr> <td>Ms Eleanor Howick</td> <td>5.01, 5.02, 5.04, 5.05, 5.11, 5.12, 5.14</td> </tr> <tr> <td>Mr Darrin Jack</td> <td>5.41, 5.42, 5.43, 5.44</td> </tr> <tr> <td>Mr Graeme Jonas</td> <td>5.05(d)(ii), 5.70</td> </tr> <tr> <td>Mr Keith Jones</td> <td>5.82, 5.84, 5.85, 5.86, 5.88(a)(c)(e), 5.89(a)(c)(e)(f)(i)(l), 5.90, 5.97</td> </tr> <tr> <td>Dr Annette Koo</td> <td>5.68, 5.69</td> </tr> <tr> <td>Dr Jeremy Lovell-Smith</td> <td>5.35</td> </tr> <tr> <td>Mr Ross Mason</td> <td>5.35, 5.61(c)</td> </tr> <tr> <td>Mr Greg Reid</td> <td>5.21, 5.31, 5.32, 5.33</td> </tr> <tr> <td>Dr Peter Saunders</td> <td>5.61, 5.82(a), 5.90(a)(c)</td> </tr> <tr> <td>Dr Francois Shindo</td> <td>5.65, 5.66, 5.67</td> </tr> <tr> <td>Mr Tom Stewart</td> <td>5.84(a)(c), 5.85(a), 5.88(b)(d)(e)(f))(power frequencies only), 5.89(b)(d)(e)(f)(g)(h)(l) (power frequencies only), 5.90(c)</td> </tr> <tr> <td>Dr Chris Sutton</td> <td>5.21, 5.31, 5.32, 5.33, 5.41, 5.42, 5.43, 5.44</td> </tr> <tr> <td>Mr Neil Swift</td> <td>5.65, 5.66, 5.67</td> </tr> <tr> <td>Dr David Rodney White</td> <td>5.35, 5.61, 5.82(a), 5.90(a)(c)</td> </tr> <tr> <td>Mr Chris Young</td> <td>5.01, 5.02, 5.04, 5.05, 5.11, 5.12, 5.14</td> </tr> </table>	Dr Tim Armstrong	5.91, 5.92(a)(c), 5.93(a)	Dr Laurie Christian	5.82, 5.85(d), 5.86, 5.87, 5.88, 5.89, 5.90, 5.92(b), 5.93(b), 5.97	Dr Mark Clarkson	5.41, 5.42, 5.43, 5.44	Mr David Cochrane	5.05(d)(ii), 5.70	Dr Adam Dunford	5.91(a)(c)(d)(g), 5.92(a)(c), 5.93(a)	Dr Murray Early	5.82, 5.86, 5.87, 5.88, 5.89(a)(b)(c)(d)(i), 5.90(a)(f)(g), 5.92(b), 5.93(b), 5.97	Mr Hamish Edgar	5.61	Dr Lucy Forde	5.02, 5.04, 5.11(f)(i)(n), 5.12, 5.14	Dr Blair Hall	5.93(b), 5.95	Mr John Hamlin	5.65, 5.66, 5.67, 5.68, 5.69	Ms Eleanor Howick	5.01, 5.02, 5.04, 5.05, 5.11, 5.12, 5.14	Mr Darrin Jack	5.41, 5.42, 5.43, 5.44	Mr Graeme Jonas	5.05(d)(ii), 5.70	Mr Keith Jones	5.82, 5.84, 5.85, 5.86, 5.88(a)(c)(e), 5.89(a)(c)(e)(f)(i)(l), 5.90, 5.97	Dr Annette Koo	5.68, 5.69	Dr Jeremy Lovell-Smith	5.35	Mr Ross Mason	5.35, 5.61(c)	Mr Greg Reid	5.21, 5.31, 5.32, 5.33	Dr Peter Saunders	5.61, 5.82(a), 5.90(a)(c)	Dr Francois Shindo	5.65, 5.66, 5.67	Mr Tom Stewart	5.84(a)(c), 5.85(a), 5.88(b)(d)(e)(f))(power frequencies only), 5.89(b)(d)(e)(f)(g)(h)(l) (power frequencies only), 5.90(c)	Dr Chris Sutton	5.21, 5.31, 5.32, 5.33, 5.41, 5.42, 5.43, 5.44	Mr Neil Swift	5.65, 5.66, 5.67	Dr David Rodney White	5.35, 5.61, 5.82(a), 5.90(a)(c)	Mr Chris Young	5.01, 5.02, 5.04, 5.05, 5.11, 5.12, 5.14
Dr Tim Armstrong	5.91, 5.92(a)(c), 5.93(a)																																																		
Dr Laurie Christian	5.82, 5.85(d), 5.86, 5.87, 5.88, 5.89, 5.90, 5.92(b), 5.93(b), 5.97																																																		
Dr Mark Clarkson	5.41, 5.42, 5.43, 5.44																																																		
Mr David Cochrane	5.05(d)(ii), 5.70																																																		
Dr Adam Dunford	5.91(a)(c)(d)(g), 5.92(a)(c), 5.93(a)																																																		
Dr Murray Early	5.82, 5.86, 5.87, 5.88, 5.89(a)(b)(c)(d)(i), 5.90(a)(f)(g), 5.92(b), 5.93(b), 5.97																																																		
Mr Hamish Edgar	5.61																																																		
Dr Lucy Forde	5.02, 5.04, 5.11(f)(i)(n), 5.12, 5.14																																																		
Dr Blair Hall	5.93(b), 5.95																																																		
Mr John Hamlin	5.65, 5.66, 5.67, 5.68, 5.69																																																		
Ms Eleanor Howick	5.01, 5.02, 5.04, 5.05, 5.11, 5.12, 5.14																																																		
Mr Darrin Jack	5.41, 5.42, 5.43, 5.44																																																		
Mr Graeme Jonas	5.05(d)(ii), 5.70																																																		
Mr Keith Jones	5.82, 5.84, 5.85, 5.86, 5.88(a)(c)(e), 5.89(a)(c)(e)(f)(i)(l), 5.90, 5.97																																																		
Dr Annette Koo	5.68, 5.69																																																		
Dr Jeremy Lovell-Smith	5.35																																																		
Mr Ross Mason	5.35, 5.61(c)																																																		
Mr Greg Reid	5.21, 5.31, 5.32, 5.33																																																		
Dr Peter Saunders	5.61, 5.82(a), 5.90(a)(c)																																																		
Dr Francois Shindo	5.65, 5.66, 5.67																																																		
Mr Tom Stewart	5.84(a)(c), 5.85(a), 5.88(b)(d)(e)(f))(power frequencies only), 5.89(b)(d)(e)(f)(g)(h)(l) (power frequencies only), 5.90(c)																																																		
Dr Chris Sutton	5.21, 5.31, 5.32, 5.33, 5.41, 5.42, 5.43, 5.44																																																		
Mr Neil Swift	5.65, 5.66, 5.67																																																		
Dr David Rodney White	5.35, 5.61, 5.82(a), 5.90(a)(c)																																																		
Mr Chris Young	5.01, 5.02, 5.04, 5.05, 5.11, 5.12, 5.14																																																		

Authorised: General Manager <i>P. Bam</i>	Issue 46	Date: 28/11/17	Page 2 of 26
--	----------	----------------	--------------

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

The uncertainty of measurement is reported as an expanded uncertainty having a level of confidence of at least 95%, unless stated otherwise. Calibrations are generally performed at the premises of the accredited laboratory, although some may be carried out in the field and some at customer premises.

Measurand	Conditions	Least uncertainty of measurement
-----------	------------	----------------------------------

5.01 Engineers' Limit Gauges

(a) Plain plug, ring and gap gauges. Taper plug and ring gauges.

Setting plug gauges by comparison with gauge blocks

Mean diameter	0.5 mm to 25 mm	Q(130, 1.4L) nm, L in mm
Mean diameter	25 mm to 300 mm	Q(95, 1.8L) nm, L in mm

Setting ring gauges by comparison with gauge blocks

Mean diameter	1 mm to 300 mm	Q(95, 1.8L) nm, L in mm
---------------	----------------	-------------------------

Where $Q(a, b) = \sqrt{a^2 + b^2}$

(e) Position and receiver gauges involving both linear and angular measurements.

Lobster tail gauges	54 mm to 60 mm	0.01 mm
---------------------	----------------	---------

(g) Other gauges involving measurements similar to those under (a) and including depth gauges, height gauges and gauges involving plane coordinated position of holes and spigots.

Step gauge face spacing by comparison with end standards on CMM

90 mm to 700 mm	Q(0.7, 1.2 x 10 ⁻³ L) µm, L in mm
-----------------	--

2D CMM artefacts (ball plate centre coordinates) by comparison with end standards on CMM


Side length between 100 mm and 600 mm	Q(0.9, 1.3 x 10 ⁻³ L) µm, L in mm
---------------------------------------	--

5.02 Jigs, Fixtures, Cutting Tools and Components

Measurement of components/objects on CMM

Error of indicated size	1 mm to 800 mm	(1.2 + L/400) µm, L in mm
-------------------------	----------------	---------------------------

Authorised:
 General Manager



Issue 46

Date: 28/11/17

Page 3 of 26

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION Accreditation No 1

5.04 Machine Tools

(b) Practical tests including:

Levelling of dynamic weigh station sites by measurement of deviation from a horizontal plane (calibration carried out on site)

Deviation in height	Horizontal range 1.8 m to 60 m	0.3 mm
---------------------	-----------------------------------	--------

5.05 Geometric Form

(b) Roundness

Variability in roundness	Range of diameters 0 µm to 400 µm	Q(0.14, 0.05R) µm, R in µm
--------------------------	--------------------------------------	----------------------------

(d) Flatness of Optical Flat, Parallelism, Wedge Angle of Optical Wedge or Flat

i) Length section

Parallelism	Range of diameters 0 µm to 10 µm	10 mm to 35 mm 0.08 µm
-------------	-------------------------------------	---------------------------

Flatness	Range of diameters 0 µm to 2.5 µm	10 mm to 35 mm 0.06 µm
----------	--------------------------------------	---------------------------

ii) Photometry section

Flatness of optical flats, one-axis or whole surface


Up to 150 mm diameter (excluding 10 mm closest to rim)	22 nm
Up to 250 mm diameter (excluding 10 mm closest to rim)	33 nm
For outer 10 mm	110 nm

5.11 Working Standards of Length and Angle

(a) Gauge blocks and accessories

Measurement of central length

By interferometry	0.5 mm to 103 mm	Q(17, 0.15L) nm, L in mm
-------------------	------------------	--------------------------

Authorised: General Manager 	Issue 46	Date: 28/11/17	Page 4 of 26
--	----------	----------------	--------------

Schedule to


CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

By comparison	0.1 mm to 103 mm	Q(36, 1.4L) nm, <i>L</i> in mm
Measurement of variation in length		Q(30, 0.35L) nm, <i>L</i> in mm
(b) Length bars and accessories		
Measurement of central length and variation in length		
Long gauge blocks by comparison with gauge blocks using the Horizontal Federal		
100 mm to 300 mm		Q(91, 1.3L) nm, <i>L</i> in mm
Measurement of variation in length		Q(34, 0.35L) nm, <i>L</i> in mm
Long gauge blocks by comparison with gauge blocks using the LBC		
100 mm to 1500 mm		Q(370, 0.48L) nm, <i>L</i> in mm
Measurement of variation in length		100 nm
(f) Precision linear scales		
Engineer or machinist scale-line spacing		
0.1 m to 4 m		Q(10, 8.2L) μ m, <i>L</i> in m
(h) Precision graticules including stage micrometers and haemocytometer counting chambers		
1 μ m to 10 mm		0.5 μ m
(i) Surveying tapes and petroleum dip tapes		
4 m to 50 m		Q(10, 10.5L) μ m, <i>L</i> in m
Surveyor levelling rods		
0.5 m to 3 m		Q(10, 10L) μ m, <i>L</i> in m
(n) Geodetic Baselines (calibrations carried out on site)		
Interval distances	2 m to 1500 m	Q(0.3, 0.6 x 10 ⁻³ L) mm, <i>L</i> in m

Authorised:
 General Manager



Issue 46

Date: 28/11/17


Page 5 of 26

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION Accreditation No 1

5.12	Precision Measuring Instruments	
	(a) Length measuring machines	
	Electronic distance measuring machines (EDMs)	
	Error of indicated displacement	1 m to 206 m Q(0.13, 7 x 10⁻⁴L) mm, L in m
	Error of indicated frequency	5 MHz to 100 MHz 0.16 x 10⁻⁶L x frequency
	Error of prism constant	0.26 mm
5.14	Laser Frequency	
	(a) Stabilised lasers of the mise en pratique	
	Absolute frequency	473 612 GHz 25 kHz
	(b) Other stabilised lasers	
	Absolute frequency	473 612 GHz 0.2 MHz
5.21	Masses	
	(a) Examination of laboratory standards of mass	
	(b) Examination of industrial standards of mass	
	(c) Determination of the mass of solid objects	
	1 mg to 100 mg	0.4 µg to 0.7 µg
	0.1 g to 1 g	0.7 µg to 1.6 µg
	1 g to 10 g	1.6 µg to 4 µg
	10 g to 100 g	4 µg to 8 µg
	0.1 kg to 1 kg	8 µg to 40 µg
	1 kg to 10 kg	1.1 x 10 ⁻⁷
	10 kg to 20 kg	1.6 x 10 ⁻⁷
	20 kg to 300 kg	1.5 x 10 ⁻⁶
	300 kg to 1000 kg	10 g to 16 g
5.31	Volumetric Equipment	
	(a) Examination of laboratory volumetric glassware including examination for compliance with the Class A or Class B requirements of the relevant national or international	

Authorised: General Manager 	Issue 46	Date: 28/11/17	Page 6 of 26
--	----------	----------------	--------------

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

standards

0.02 mL to 2 mL 0.0002 mL

(b) Examination of other types of volumetric apparatus

0.002 L to 50 L 0.01 %

5.32 Density

(a) Density of solids

1400 kg/m³ to 3000 kg/m³ 1.0 x 10⁻⁵

7800 kg/m³ to 8200 kg/m³ 1.5 x 10⁻⁵

(b) Density of liquids

600 kg/m³ to 2000 kg/m³ 2.0 x 10⁻⁵

5.33 Hydrometers

(a) Density hydrometers

(b) Specific gravity hydrometers

(c) Brix hydrometers

(d) Proof spirit hydrometers

600 kg/m³ to 2000 kg/m³ 2.0 x 10⁻⁵

5.35 Hygrometry

(a) Humidity measuring devices

i) Dew point hygrometers

-70 °C to 0 °C 0.2 °C to 0.06 °C

0 °C to 40 °C 0.06 °C

40 °C to 70 °C 0.06 °C to 0.12 °C

ii) Relative humidity hygrometers

10 % to 95 % 0.006 x *h* %

(Temperature between 0 °C and 70 °C) *h* is relative humidity expressed as a percentage, that is % rh

Authorised: General Manager <i>P. Bam</i>	Issue 46	Date: 28/11/17	Page 7 of 26
--	----------	----------------	--------------

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

iii) Dry bulb temperature

0 °C to 70 °C

0.1 °C

5.41 Barometers

(a) Aneroid barometers (including digital barometers)

50 kPa to 90 kPa

2.0×10^{-5}

90 kPa to 110 kPa

1.0×10^{-5}

110 kPa to 130 kPa

2.0×10^{-5}

5.42 Differential Pressure Measuring Devices (including Manometers)

(a) Diaphragm types

(b) Liquid column types, inclined and vertical

(c) Other types

1 Pa to 10000 Pa

$(6 \times 10^{-3} + 4.5 \times 10^{-5} p)$
 Pa, p in Pa

5.43 Pressure Gauge Testers and Pressure Balances

i) Absolute pressure – gas medium

8 kPa to 550 kPa

2×10^{-5}

550 kPa to 7000 kPa

6×10^{-5}

ii) Gauge pressure – gas medium

-100 kPa to -10 kPa

1×10^{-4}

-10 kPa to -1 kPa

200 mPa to 100 mPa,
 decreasing linearly
 100 mPa to 160 mPa,
 increasing linearly

1 kPa to 8 kPa

8 kPa to 550 kPa

2×10^{-5}

550 to 11000 kPa

6×10^{-5}

iii) Gauge pressure – liquid medium


0.1 MPa to 17 MPa

$(1 \times 10^{-4} + 6.6 \times 10^{-5} p)$
 MPa (p in MPa)

17 MPa to 280 MPa

$(6.6 \times 10^{-5} p + 7 \times 10^{-7} p^2)$
 MPa (p in MPa)

Authorised:
 General Manager



Issue 46

Date: 28/11/17

Page 8 of 26

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

5.44 Pressure and Vacuum Measurement

- (a) Pressure gauges
- (b) Vacuum gauges
- (c) Pressure transducers
- (d) Pressure recorders

i) Absolute pressure – gas medium

8 kPa to 90 kPa	2×10^{-5}
90 kPa to 110 kPa	1×10^{-5}
110 kPa to 550 kPa	2×10^{-5}
550 kPa to 7000 kPa	6×10^{-5}

ii) Gauge pressure – gas medium

-96 kPa to 8 kPa	0.0031 kPa
8 kPa to 90 kPa	2×10^{-5}
90 kPa to 110 kPa	1×10^{-5}
110 kPa to 550 kPa	2×10^{-5}
550 to 11000 kPa	6×10^{-5}

iii) Absolute pressure – liquid medium

0.3 MPa to 17 MPa	$(1 \times 10^{-4} + 6.6 \times 10^{-5}p)$ MPa (p in MPa)
17 MPa to 280 MPa	$(6.6 \times 10^{-5}p + 7 \times 10^{-7}p^2)$ MPa (p in MPa)


iv) Gauge pressure – liquid medium

0.2 MPa to 17 MPa	$(1 \times 10^{-4} + 6.6 \times 10^{-5}p)$ MPa (p in MPa)
17 MPa to 280 MPa	$(6.6 \times 10^{-5}p + 7 \times 10^{-7}p^2)$ MPa (p in MPa)

5.61 Temperature Measuring Equipment

- (c) Platinum (and other metallic) resistance thermometers
- i) Contact thermometers, including Standard PRTs at the following fixed points

Argon triple point (-189.3442 °C)	1 mK
-----------------------------------	------

Authorised: General Manager 	Issue 46	Date: 28/11/17	Page 9 of 26
--	----------	----------------	--------------

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

Mercury triple point (-38.8344 °C)	0.4 mK
Water triple point (0.01 °C)	0.1 mK
Gallium melting point (29.7646 °C)	0.19 mK
Indium freezing point (156.5985 °C)	0.56 mK
Tin freezing point (231.928 °C)	0.85 mK
Zinc freezing point (419.527 °C)	1.9 mK
Aluminium freezing point (660.323 °C)	10 mK

ii) Industrial PRTs and direct reading thermometers

-190 °C to <0 °C	(2.4 - 0.005 x t) mK, t in °C
0 °C to 200 °C	(2.4 + 0.008 x t) mK, t in °C
>200 °C to 550 °C	(4.0 + 0.03 x (t - 200)) mK, t in °C

5.65 Photometers and Radiometers

(a) Photometers

10 lux to 3000 lux	0.8 %
--------------------	-------

(b) Illuminance meters

0.005 lux to 10 lux	3 %
10 lux to 3000 lux	0.8 %

(c) Luminance meters

2 cd/m ² to 800 cd/m ²	1.4 %
800 cd/m ² to 27000 cd/m ²	7 %
27000 cd/m ² to 33000 cd/m ²	11 %

(d) UV meters

240 nm to 250 nm	1.4 %
250 nm to 280 nm	1.4 % to 0.39 %, decreases linearly with wavelength
280 nm to <420 nm	0.39 %


(g) Laser power meters

Laser lines from 450 nm to 800 nm	0.022 %
-----------------------------------	---------

(h) Detector spectral responsivity measurement

240 nm to 250 nm	1.4 %
250 nm to 280 nm	1.4 % to 0.39 %, decreases linearly with wavelength

Authorised:
General Manager



Issue 46

Date: 28/11/17

Page 10 of 26

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

280 nm to <420 nm	0.39 %
420 nm to 680 nm	0.06 %
680 nm to 800 nm	0.08 %
800 nm to 950 nm	0.14 % to 0.16 %, varies with wavelength

Discrete wavelengths

Laser lines from 450 nm to 800 nm	0.022 %
-----------------------------------	---------

5.66 Lamps, LEDs, Lasers and Other Light Sources

Calibrations within 5.66 may be offered in the field as well as in the laboratory. An increase in uncertainty due to environmental conditions and other influence variables present in the field may need to be applied.

(a) Lamps: luminous intensity

10 cd to 5000 cd	0.8 %
------------------	-------

(e) Illuminance

0.005 lux to 3000 lux	3 %
-----------------------	-----

(f) General sources: spectral irradiance

250 nm to 350 nm	0.001 W/(m ² .nm) to 0.5 W/(m ² .nm)	2.6 % to 1.6 %
350 nm to 850 nm	0.001 W/(m ² .nm) to 0.5 W/(m ² .nm)	1.6 % to 1.4 %

(h) Photoluminescent materials

from 0.5 mcd/m ²	0.5 mcd/m ² or 15 %, whichever is greater
-----------------------------	--


5.67 Colour of Light Sources and Colorimeters

Calibrations within 5.67 may be offered in the field as well as in the laboratory. An increase in uncertainty due to environmental conditions and other influence variables present in the field may need to be applied.

(a) General sources:

Colour emitted in CIE x, y colour space	0.0005 to 0.005 in x and y, varies with measurand
---	---

Authorised:
General Manager



Issue 46

Date: 28/11/17

Page 11 of 26

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

Colour emitted in CIE u, v colour space 0.0007 in u and v

(d) Lamps:

Correlated colour temperature 2700 K to 3000 K 5.4 K to 6.3 K, varies with measurand

5.68 Optical Properties of Materials: Spectral

(a) Regular transmittance and absorbance (bandwidth 1 nm to 3 nm)

240 nm to 380 nm	0.01 to 1.0	0.5 % of value
380 nm to 1000 nm	0.0001 to 0.01	0.00005
380 nm to 1000 nm	0.01 to 0.1	0.00005 to 0.0001
		varies with transmittance
380 nm to 1000 nm	0.1 to 1.0	0.1 % of value

(b) Wavelength calibration filters

240 nm to 800 nm	0.13 nm
800 nm to 1100 nm	0.13 nm to 0.25 nm

(c) Diffuse transmittance

300 nm to 400 nm	0.005 to 0.0002 or 5 % of value whichever is greater
400 nm to 1000 nm	0.0002 or 5 % of value whichever is greater

(d) Diffuse reflectance in 0/d and 6/d geometries

360 nm to 820 nm	0.016 to 0.9	0.008 to 0.0036, varies with wavelength
360 nm to 820 nm	0.9 to 1.0	0.4 % of value

(e) Specular reflectance at normal incidence


280 nm to 800 nm	0.05 to 1	1 % of value
------------------	-----------	--------------

5.69 Optical Properties of Materials: Spectrally integrated

(a) Luminous transmittance

Spectrally flat materials	0.3 % of value
General materials	5 % of value

Authorised:
 General Manager



Issue 46

Date: 28/11/17

Page 12 of 26

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION Accreditation No 1

(b)	Luminous reflectance	
	General materials	5 % of value
(c)	Colour transmitted, x, y, Y or L*a*b*	
	In x and y	0.005
	Luminous transmittance Y for (0.1 < Y < 1)	5 % of value
(d)	Colour of surfaces, x, y, Y or L*a*b*	
	In x and y	0.003
	Luminance factor Y for (0.1 < Y < 1)	5 % of value
(e)	Retroreflectors: CIL value	
	Coefficient of luminous intensity	5 %

5.70 Optical Instruments

(a)	Focal length	0.03 mm
(b)	Image plane principal point and nodal points	0.03 mm
	Nodal points	0.05 mm
(c)	Field of view	0.5 degrees

5.82 Resistors, Resistance Boxes and Potential Dividers

(a)	Precision resistors, resistance boxes and conductance boxes	
	0.1 Ω to 1 Ω (Current ≤ 100 mA)	0.2 μΩ/Ω
	1 Ω to 10 kΩ (Power dissipation ≤ 10 mW)	0.12 μΩ/Ω
	10 mΩ to 1000 mΩ (Current ≤ 1A)	25 μΩ/Ω
	0.1 mΩ to 1000 mΩ (Current = 1 A to 875 A)	63 R ^{-0.35} μΩ/Ω, R in mΩ values range from 141 μΩ/Ω to 6 μΩ/Ω
	0.01 MΩ to 1 MΩ (Applied voltages = 5 V to 100 V)	0.7 μΩ/Ω

Authorised: General Manager	Issue 46	Date: 28/11/17	Page 13 of 26
--------------------------------	----------	----------------	---------------

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION Accreditation No 1

0.001 GΩ to 1 GΩ
 (Applied voltages = 5 V to 100 V) (0.7 + 27 R - 20 R³) μΩ/Ω, R in GΩ, values range from 0.7 μΩ/Ω to 8 μΩ/Ω

1 MΩ to 5 TΩ
 (Applied voltages = 100 V to 1000 V) (35 + 6.9 × 10⁻¹¹ R² + 9.4 μΩ/Ω × 10⁻⁴ R) μΩ/Ω, R in MΩ, values range from 35 μΩ/Ω to 6460 μΩ/Ω

0 MΩ to 1 MΩ
 (frequency, f = 40 Hz to 2 kHz) (2000/f + 19 R) μΩ, f in Hz, R in Ω, values range from 1 μΩ to 19 Ω

0 Ω to 400 Ω
 (frequency, f = 0.01 Hz to 100 Hz) (6 + 0.3 R) μΩ, R in Ω, values range from 6 μΩ to 126 μΩ

(b) Volt ratio boxes and potential dividers

1 V/V to 1000 V/V
 (Input voltage ≤ 1100 V, output voltage ≥ 1 V) 0.4 × 10⁻⁶

0 kV to 50 kV 3 mV/V

(c) DC shunts

0.1 mΩ to 1 Ω
 (Applied current 1 A to 875 A)
 (Applied voltage 10 mV to 1 V) 63 R^{0.35} μΩ/Ω, R in mΩ values range from 141 μΩ/Ω to 6 μΩ/Ω

(d) AC shunts

0 Ω to 100 Ω
 (frequency, f = 40 Hz to 2 kHz) (2000/f + 19R) μΩ, f in Hz, R in Ω, values range from 1 μΩ to 1900 μΩ


0.2 A to 100 A
 (frequency, f = 47 Hz to 75 Hz) 25 μΩ/Ω

5.84 Capacitors

(a) Precision capacitors

0 μF to 100 μF
 (frequency, f = 40 Hz to 2 kHz) (0.2/f + 22C) pF, f in Hz, C in μF, values range from 0.0001 pF to 2200 pF

Dissipation factor

Authorised: General Manager 	Issue 46	Date: 28/11/17	Page 14 of 26
--	----------	----------------	---------------

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

0 to 0.2
(frequency, $f = 40$ Hz to 2 kHz)
(capacitance, $C = 0.5$ pF to 100 μ F) (0.000027 + 0.00027/C) C in pF,
values range from 0.00057 to
0.000027

(c) Capacitance potential dividers

1 kV rms to 35 kV rms 1 mV/V
(frequency, $f = 50$ Hz to 3 kHz)

5.85 Inductors and Transformers

(a) Inductors, self and mutual

0 H to 100 H (0.2/f + 14L)H μ H, f in Hz
 L in H, values range from
0.0001 μ H to 1400 μ H
(frequency, $f = 40$ Hz to 2 kHz)

Equivalent series resistance
0 Ω to 1 M Ω (2000/f + 19R) $\mu\Omega$, f in Hz,
 R in Ω , values range from 1
 $\mu\Omega$ to 19 Ω
(frequency, $f = 40$ Hz to 2 kHz)

(d) Current transformers: protection and measurement

Primary currents 1 A to 4000 A, ratios 0.2 A/A to 4000 A/A

Ratio error -25 % to 25 % 0.0010 % to 0.13 %
Phase error -36 crad to 36 crad 0.0010 crad to 0.18 crad
(frequency, $f = 50$ Hz; secondary currents 1 A, 5 A)

5.86 Voltage Standards and Current Standards

(b) Electronic emf reference devices


1 V 0.1 μ V
1.018 V 0.1 μ V
10 V 1.5 μ V

5.87 Transfer Instruments (AC/DC)

0.002 V to 0.6 V 11 μ V/V to 321 μ V/V
> 0.6 V to 6 V 6 μ V/V to 77 μ V/V
> 6 V to 1000 V 9 μ V/V to 76 μ V/V
(frequency, $f = 10$ Hz to 1 MHz)

1 V and 3 V 0.16 mV/V to 2.6 mV/V
(frequency, $f = 1$ MHz to 100 MHz)

Authorised:
General Manager



Issue 46

Date: 28/11/17

Page 15 of 26

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

0.1 mA to 0.01 A
 (frequency, $f = 40$ Hz to 2 kHz) 15 μ A/A to 38 μ A/A

0.01 A to 20 A
 (frequency, $f = 40$ Hz to 100 kHz) 15 μ A/A to 70 μ A/A

5.88 Calibrators for Instrumentation

(a) DC voltage

0 V to 12 V (0.05 + 0.15 U) μ V, U in V, values range from 0.05 μ V to 1.85 μ V

12 V to 1100 V 0.5 μ V/V

(b) AC voltage

0.002 V to 1000 V
 (frequency, $f = 10$ Hz to 1 MHz) 6 μ V/V to 650 μ V/V

1 V and 3 V
 (frequency, $f = 1$ MHz to 100 MHz) 0.3 mV/V to 8 mV/V

(c) DC current

1x10⁻¹¹ A to 1x10⁻⁵ A values range from 5 μ A/A to 560 μ A/A

0.01 mA to 0.1 mA 5 μ A/A

>0.1 mA to 1 A 5 μ A/A

>1 A to 20 A $5 I^{0.43}$ μ A/A, I in A, values range from 5 μ A/A to 18 μ A/A

20 A to 1000 A $5 I^{0.43}$ μ A/A, I in A, values range from 18 μ A/A to 97 μ A/A

(d) AC current

0.1 mA to 2 A
 (frequency, $f = 40$ Hz to 2 kHz) 35 μ A/A to 170 μ A/A

0.01 A to 100 A
 (frequency, $f = 47$ Hz to 75 Hz) 25 μ A/A


(e) Resistance

0 Ω to 10 Ω 40 $\mu\Omega$

0.01 k Ω to 1 M Ω 3 $\mu\Omega/\Omega$

1 M Ω to 100 M Ω (2 + $R^{0.8}$) $\mu\Omega/\Omega$, R in M Ω , values range from 3 $\mu\Omega/\Omega$ to

Authorised:
 General Manager



Issue 46

Date: 28/11/17

Page 16 of 26

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

42 $\mu\Omega/\Omega$

(f) AC power sources

Same as 5.89 (e) and (f)

5.89 Indicating Instruments and Recording Instruments

(a) DC voltmeters

0 V to 0.001 V

0.001 V to 12 V

12 V to 1100 V

0.05 μV

(0.05 + 0.15 U) μV , U in V, values range from 0.05 μV to 1.85 μV

0.5 $\mu\text{V/V}$

(b) AC voltmeters

0.002 V to 1000 V

(frequency, $f = 10$ Hz to 1 MHz)

1 V and 3 V

(frequency, $f = 1$ MHz to 100 MHz)

9 $\mu\text{V/V}$ to 862 $\mu\text{V/V}$

0.3 mV/V to 8 mV/V

(c) DC ammeters

1x10⁻¹¹ A to 1x10⁻⁵ A

0.01 mA to 0.1 mA

>0.1 mA to 1 A

>1 A to 20 A

20 A to 875 A

values range from 5 $\mu\text{A/A}$ to 560 $\mu\text{A/A}$

5 $\mu\text{A/A}$

5 $\mu\text{A/A}$

5 $I^{0.43}$ $\mu\text{A/A}$, I in A, values range from 5 $\mu\text{A/A}$ to 18 $\mu\text{A/A}$

5 $I^{0.43}$ $\mu\text{A/A}$, I in A

values range from 18 $\mu\text{A/A}$ to 92 $\mu\text{A/A}$

(d) AC ammeters

0.1 mA to 2 A

(frequency, $f = 40$ Hz to 2 kHz)

0.2 A to 100 A

(frequency, $f = 47$ Hz to 75 Hz)

60 $\mu\text{A/A}$ to 140 $\mu\text{A/A}$


25 $\mu\text{A/A}$

(e) Wattmeters

Conditions

Voltage 60 V to 240 V, current 0.01 A to 100 A, frequency 45 Hz to 75 Hz, and PF 1 to 0, inductive or capacitive

Authorised:
 General Manager



Issue 46

Date: 28/11/17

Page 17 of 26

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

Single phase
 0 W to 24000 W (40 μ W/VA + 6(1-PF)), values range from 40 μ W/VA to 46 μ W/VA

Three phase
 0 W to 72000 W (40 μ W/VA + 6(1-PF)), values range from 40 μ W/VA to 46 μ W/VA

(The range and uncertainties for star and delta are the same as for single-phase)

(f) Varmeters

Conditions
 Voltage 60 V to 240 V, current 0.01 A to 100 A, frequency 45 Hz to 75 Hz, and QF 1 to 0, inductive or capacitive

0 W to 24000 W (40 μ Var/VA + 90QF), values range from 40 μ Var/VA to 130 μ Var/VA

0 W to 72000 W (40 μ Var/VA + 90QF), values range from 40 μ Var/VA to 130 μ Var/VA

(g) Phase angle indicators (source or meter)

Conditions
 Current 0.01 A to 100 A, frequency 45 Hz to 75 Hz, Voltage 0.7 V to 7 V, 42 V to 240 V

-3.14 rad to 3.14 rad 40 μ rad

(h) Power factor meters


Same conditions, range and uncertainty as 5.89 (g) above

(i) Ohmmeters

0.1 m Ω to 1000 m Ω
 (applied current 875 A to 1 A) 63R^{0.35} $\mu\Omega$, R in m Ω , values range from 141 $\mu\Omega$ to 6 $\mu\Omega$

0.1 Ω to 1 Ω
 (applied current \leq 100 mA) 0.2 $\mu\Omega/\Omega$

1 Ω to 10 k Ω
 10 k Ω to 1 G Ω 0.12 $\mu\Omega/\Omega$
 (1 + 27R - 20R³) $\mu\Omega/\Omega$, R in G Ω , values range from 1 $\mu\Omega/\Omega$ to 8 $\mu\Omega/\Omega$

Authorised: General Manager		Issue 46	Date: 28/11/17	Page 18 of 26
--------------------------------	---	----------	----------------	---------------

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION Accreditation No 1

1 GΩ to 100 GΩ	(-0.07R ² + 22R - 15) μΩ/Ω, R in GΩ, values range from 6.9 μΩ/Ω to 1485 μΩ/Ω
100 GΩ to 1200 GΩ	(1300R + 2.2R) μΩ/Ω, R in GΩ, values range from 1520 μΩ/Ω to 3940 μΩ/Ω

(k) Galvanometers and null detectors
 Same range and least uncertainties from 5.89 (a) DC voltmeters

(l) Energy meters
 Same as 5.89 (e) and (f)

5.90 Bridges, Potentiometers and Test Sets

(a) DC bridges
 Same as 5.89 (i) Ohmmeters above
 DC Bridges for thermometry

Resistance ratio
 0 Ω/Ω to 13 Ω/Ω 2 x 10⁻⁸

(b) DC potentiometers
 Same as 5.89 (a) DC Voltmeters above


(c) AC bridges

0 Ω to 1 MΩ	(2000/f + 19R) μΩ, f in Hz, R in Ω, values range from 1 μΩ to 19 Ω
0 μF to 100 μF	(0.2/f + 22C) pF, f in Hz, C in μF, values range from 0.0001 pF to 2200 pF
0 H to 1 H	(0.2/f + 14L) H μH, f in Hz, L in H, values range from 0.0001 μH to 14 μH

(frequency, f = 40 Hz to 2 kHz)

AC Bridges for thermometry

0 Ω/Ω to 13 Ω/Ω (frequency, f = 0 Hz to 100 Hz)	2 x 10 ⁻⁸
--	----------------------

Authorised: General Manager 	Issue 46	Date: 28/11/17	Page 19 of 26
--	----------	----------------	---------------

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

(f) Current transformer testing sets

Ratio/Phase
 (frequencies in the range 45 Hz to 65 Hz)

Ratio error	$\pm (0 \text{ to } 0.002)$	5.0×10^{-7} to 1.0×10^{-6}
Ratio error	$\pm (0.002 \text{ to } 0.02)$	2.0×10^{-6} to 8.0×10^{-6}
Ratio error	$\pm (0.02 \text{ to } 0.2)$	2.0×10^{-5} to 8.0×10^{-5}
Phase error	$\pm 0 \text{ rad to } 0.002 \text{ rad}$	5.0×10^{-7} rad to 1.0×10^{-6} rad
Phase error	$\pm 0.002 \text{ rad to } 0.02 \text{ rad}$	5.0×10^{-6} rad to 9.0×10^{-6} rad
Phase error	$\pm 0.02 \text{ rad to } 0.2 \text{ rad}$	5.0×10^{-5} rad to 9.0×10^{-5} rad

(g) Voltage transformer testing sets

Same as 5.90 (f)

5.91 Frequency Measurement and Time Measurement

Time and frequency Least Uncertainties relate only to the reference measuring systems. These uncertainties do not contain any contribution from the instrument under calibration.

(a) Frequency meters

1 Hz to 40 GHz	1×10^{-10}
0.001 Hz to 1 Hz (period)	1 ns

(c) Counters

1 Hz to 40 GHz	1×10^{-10}
0.001 Hz to 1 Hz (period)	1 ns

(d) Time interval meters

10 ns to 86400 s	2 ns or 2×10^{-13} , whichever is greatest
------------------	---

(e) Clocks and watches


5 ns to 1 s	50 ns
-------------	-------

(g) Frequency standards

100 kHz to 10 MHz	2×10^{-13}
0.001 Hz to 1 Hz (period)	1 ns

(h) Time measurement

Authorised:
 General Manager



Issue 46

Date: 28/11/17

Page 20 of 26

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

Time scale difference –
 Local clock vs UTC (MSL)
 -1 s to 1 s

2 ns

Time scale difference –
 Local clock vs UTC
 -1 s to 1 s

50 ns

5.92 Waveform Measurement

(a) Frequency characteristics

1 Hz to 20 MHz
 0.001 Hz to 1 Hz (period)

1 in 10^{-10}
 1 ns

(b) Input characteristics

1 V and 3 V
 (frequency, $f = 1$ MHz to 100 MHz)

0.16 mV/V to 2.6 mV/V

Pulse risetime ($T > 5$ ns)
 (10 mV to 10 V)

0.005 μ s to 1.00×10^6 μ s

$Q(2$ ns, $0.05T)$, T in s

Pulse amplitude (pulse length > 200 μ s)
 (10 mV, 100 mV, 1 V, 10 V)

0 V to 10 V

$(30 \mu$ V + $100V_a + 420V_r)$,
 applied voltage V_a in V,
 voltmeter range V_r in V, values
 range from 34.2 μ V to 5230 μ V

(c) Timing characteristics

10 ns to 100 s
 (time difference)

2 ns

5.93 Signal Sources


(a) Frequency characteristics

1 Hz to 20 MHz
 0.001 Hz to 1 Hz (period)

1×10^{-10}
 1 ns

(b) Output characteristics

Authorised:
 General Manager



Issue 46

Date: 28/11/17

Page 21 of 26

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

1 V and 3 V (frequency, $f = 1 \text{ MHz to } 100 \text{ MHz}$) Pulse amplitude (pulse length $> 200 \mu\text{s}$) (10 mV, 100 mV, 1 V, 10 V) 0 V to 10 V Pulse risetime ($T > 5 \text{ ns}$) (10 mV to 10 V) 0.005 μs to $1.00 \times 10^6 \mu\text{s}$ RF power (absolute) 0.1 mW to 10 mW (frequency range from 10 MHz to 2.5 GHz) Reflection coefficient -1.0 to 1.0 (frequency range from 10 MHz to 2.5 GHz)	0.16 mV/V to 2.6 mV/V (30 $\mu\text{V} + 100 V_a + 420 V_r$) applied voltage V_a in V, voltmeter range V_r in V, values range from 34.2 μV to 5230 μV Q(2 ns, 0.05 T), T in s 0.0026 mW/mW 0.0050
--	---

5.95 Communications Equipment

(h) Power measuring equipment

Calibration factor Values from 0.1 to 1.2 (frequency range 10 MHz to 18 GHz, power level -15 dBm to 10 dBm)	0.004 to 0.012
--	----------------

- (i) Attenuators and amplifiers
- (j) Waveguide and coaxial components

Measurement of reflection coefficient for coaxial components in the range 30 kHz to 18 GHz

Complex reflection coefficient with magnitude between 0 and 1

Nominal magnitude (linear)	Least uncertainty of measurement (real or imaginary component)
1.0	0.002 to 0.017
0.9	0.002 to 0.015
0.8	0.002 to 0.013
0.7	0.002 to 0.011
0.6	0.003 to 0.010

Authorised: General Manager <i>P. Bone</i>	Issue 46	Date: 28/11/17	Page 22 of 26
---	----------	----------------	---------------

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION Accreditation No 1

0.5	0.003 to 0.010
0.4	0.003 to 0.009
0.3	0.003 to 0.009
0.2	0.004 to 0.008
0.1	0.004 to 0.008
0.0	0.004 to 0.008 (increases with frequency)

Reflection coefficient magnitude between 0 and 1

Nominal magnitude (linear)	Least uncertainty of measurement (magnitude)
1.0	0.002 to 0.017
0.9	0.002 to 0.015
0.8	0.002 to 0.013
0.7	0.002 to 0.011
0.6	0.003 to 0.010
0.5	0.003 to 0.010
0.4	0.003 to 0.009
0.3	0.003 to 0.009
0.2	0.004 to 0.008
0.1	0.004 to 0.008
0.0	0.004 to 0.008 (increases with frequency)


Reflection coefficient phase between -180° and 180°

Nominal magnitude (linear)	Least uncertainty of measurement (phase)
1.0	0.1° to 1.0°
0.9	0.1° to 1.0°
0.8	0.2° to 1.0°
0.7	0.2° to 1.0°
0.6	0.3° to 1.0°
0.5	0.3° to 1.1°
0.4	0.5° to 1.3°
0.3	0.6° to 1.7°
0.2	1.0° to 2.4°
0.1	2.1° to 4.7°
0.0	180° (increases with frequency)

Measurement of transmission coefficient for coaxial components in the range 30 kHz to 100 kHz

Complex transmission coefficient with magnitude between 0 and 1

Nominal magnitude (linear)	Least uncertainty of measurement (relative uncertainty of real or imaginary)

Authorised: General Manager 	Issue 46	Date: 28/11/17	Page 23 of 26
--	----------	----------------	---------------

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

	component)
1.000	0.003
0.708	0.004
0.500	0.004
0.230	0.004
0.100	0.004
0.032	0.005
0.010	0.008 to 0.006
0.003	0.018 to 0.006
0.001	0.050 to 0.009
	(decreases with increasing frequency)

Transmission coefficient magnitude between 0 and 60 dB

Nominal magnitude (log)	Least uncertainty of measurement (magnitude)
0 dB	0.03 dB
3 dB	0.03 dB
6 dB	0.03 dB
10 dB	0.03 dB
20 dB	0.04 dB
30 dB	0.05 dB to 0.04 dB
40 dB	0.07 dB to 0.05 dB
50 dB	0.15 dB to 0.06 dB
60 dB	0.42 dB to 0.08 dB
	(decreases with increasing frequency)

Transmission coefficient phase between -180° and 180°

Nominal magnitude (log)	Least uncertainty of measurement (phase)
0 dB	0.2°
3 dB	0.2°
6 dB	0.2°
10 dB	0.2°
20 dB	0.3°
30 dB	0.3°
40 dB	0.4° to 0.3°
50 dB	1.0° to 0.4°
60 dB	2.9° to 0.5°
	(decreases with increasing frequency)

Measurement of transmission coefficient for coaxial components in the range 100 kHz to 18 GHz

Authorised: General Manager <i>P. Bam</i>	Issue 46	Date: 28/11/17	Page 24 of 26
--	----------	----------------	---------------

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION

Accreditation No 1

Complex transmission coefficient with magnitude between 0 and 1

Nominal magnitude (linear)	Least uncertainty of measurement (relative uncertainty of real or imaginary component)
1.000	0.003 to 0.005
0.708	0.004 to 0.005
0.500	0.004 to 0.005
0.230	0.004 to 0.005
0.100	0.004 to 0.006
0.032	0.005 to 0.006
0.010	0.006 to 0.007
0.003	0.006 to 0.009
0.001	0.009 to 0.011

(increases with frequency)

Transmission coefficient magnitude between 0 dB and 60 dB

Nominal magnitude (log)	Least uncertainty of measurement (magnitude)
0 dB	0.03 dB to 0.04 dB
3 dB	0.03 dB to 0.04 dB
6 dB	0.03 dB to 0.05 dB
10 dB	0.03 dB to 0.05 dB
20 dB	0.04 dB to 0.05 dB
30 dB	0.05 dB
40 dB	0.05 dB to 0.06 dB
50 dB	0.06 dB to 0.08 dB
60 dB	0.08 dB to 0.10 dB


(increases with frequency)

Transmission coefficient phase between -180° and 180°

Nominal magnitude (log)	Least uncertainty of measurement (phase)
0 dB	0.2° to 0.3°
3 dB	0.2° to 0.3°
6 dB	0.2° to 0.3°
10 dB	0.2° to 0.3°
20 dB	0.3°
30 dB	0.3° to 0.4°
40 dB	0.3° to 0.4°
50 dB	0.4° to 0.5°
60 dB	0.5° to 0.6°

(increases with frequency)

Authorised:
 General Manager



Issue 46

Date: 28/11/17

Page 25 of 26

Schedule to

CERTIFICATE OF ACCREDITATION

Callaghan Innovation
 Metrology & Calibration Laboratory
SCOPE OF ACCREDITATION Accreditation No 1

5.97 High Voltage Testing


(a) Direct voltage

0 kV to 50 kV 3 mV/V

(b) Alternating voltage

1 kV rms to 35 kV rms 1 mV/V
 (frequency, $f = 50 \text{ Hz to } 3 \text{ kHz}$)

Uncontrolled copy printed from the Internet

Authorised: General Manager 	Issue 46	Date: 28/11/17	Page 26 of 26
--	----------	----------------	---------------